

MAH558

STÄUBLI

Typ/Type **XZGL-425**

600 V, CAT II

Hersteller/Producer:

Stäubli Electrical Connectors Essen GmbH

Westendstrasse 10

45143 Essen/Germany

Tel. +49 201 831 05 0

Fax +49 201 831 05 99

mail ec.essen@staubli.com

www.staubli.com/electrical



⚠ Vor jedem Gebrauch ist visuell zu prüfen, ob keine äußeren Mängel vorhanden sind (besonders an der Isolation). Bei sichtbaren Beschädigungen der Baugruppe darf diese nicht mehr verwendet werden.


Baugruppen dürfen nur innerhalb der angegebenen Bemessungsdaten verwendet werden.

Bei einer Kombination von Messzubehör aus verschiedenen Messkategorien bestimmt die niedrigste Kategorie die Messkategorie des gesamten Aufbaus.

⚠ Each time the connector is used, it should previously be inspected for external defects (particularly the insulation). In case of any visible damage, the component must no longer be used.

Assemblies must only be used within the specified ratings.

When using a combination of test accessories from different measurement categories, the lowest category determines the measurement category for the entire assembly.



Beim Zusammenstecken der beiden Tüllenteile leichten seitlichen Druck ausüben und Einrasten kontrollieren!

When fitting together the two parts of the insulator, press lightly together and check that they engage!